

Electrolyte Panel, Serum

Overview

Useful For

Identifying a suspected imbalance in electrolytes or acid/base imbalance

Method Name

KS, NAS, CL: Potentiometric, Indirect Ion-Selective Electrode HCO3: Photometric/Enzymatic AGAP: Calculated Result

NY State Available

Yes

Specimen

Specimen Type Serum

Necessary Information Patient's age and sex are required.

Specimen Required Collection Container/Tube: Preferred: Serum gel Acceptable: Red top Submission Container/Tube: Plastic vial Specimen Volume: 0.5 mL Collection Instructions: 1. Serum gel tubes should be centrifuged within 2 hours of collection. 2. Red-top tubes should be centrifuged, and the serum aliquoted into a plastic vial within 2 hours of collection.

Forms

If not ordering electronically, complete, print, and send a Renal Diagnostics Test Request (T830) with the specimen.

Specimen Minimum Volume

0.4 mL

Reject Due To

Gross	Reject
hemolysis	
Gross lipemia	ОК



Specimen Stability Information

Specimen Type	Temperature	Time	Special Container
Serum	Refrigerated	24 hours	

Clinical & Interpretive

Clinical Information

The electrolyte panel is ordered to identify electrolyte, fluid, or pH imbalance. Electrolyte concentrations are evaluated to assist in investigating conditions that cause electrolyte imbalances such as dehydration, kidney disease, lung diseases, or heart conditions. Repeat testing of the electrolyte or its components may be used to monitor the patient's response to treatment of any condition that may be causing the electrolyte, fluid or pH imbalance.

Electrolyte and acid-base imbalances can often be indicative of many acute and chronic illnesses. For this reason, the electrolyte panel is often used in the hospital and emergency settings to evaluate patients.

Reference Values

SODIUM <1 year: not established > or =1 year: 135-145 mmol/L

POTASSIUM <1 year: not established > or =1 year: 3.6-5.2 mmol/L

CHLORIDE

<1 year: not established 1-17 years: 102-112 mmol/L > or =18 years: 98-107 mmol/L

BICARBONATE

Males <1 year: not established 1-2 years: 17-25 mmol/L 3 years: 18-26 mmol/L 4-5 years: 19-27 mmol/L 6-7 years: 20-28 mmol/L 8-17 years: 21-29 mmol/L > or =18 years: 22-29 mmol/L Females <1 year: not established 1-3 years: 18-25 mmol/L 4-5 years: 19-26 mmol/L



Electrolyte Panel, Serum

6-7 years: 20-27 mmol/L 8-9 years: 21-28 mmol/L > or =10 years: 22-29 mmol/L

ANION GAP

<7 years: not established > or =7 years: 7-15

Interpretation

With an imbalance of a single electrolyte, such as sodium or potassium, repeat testing may be ordered of that particular electrolyte, can be used to monitor the imbalance until remedied. With an acid-base imbalance, blood gases may be ordered, which will measure the oxygen, carbon dioxide, and pH levels in the arterial blood. These tests assist in evaluating the acuteness of the imbalance and monitoring the response to treatment.

Cautions

No significant cautionary statements

Clinical Reference

1. Oh MS: Evaluation of renal function, water, electrolytes, and acid-base balance. <u>In</u> Henry's Clinical Diagnosis and Management by Laboratory Methods. 22nd edition. Edited by RA McPherson, MR Pincus. Philadelphia, PA: Elsevier Saunders; 2011:chap 14

2. AACC: Lab Tests Online: Access 03/22/2017. Available at https://labtestsonline.org/understanding/analytes/electrolyes

Performance

Method Description

Potassium; Chloride; Sodium:

The ion-selective electrode (ISE) module indirectly measures the electromotive force (EMF) difference between an ISE and a reference electrode. The EMF of the ISE is dependent on the ion concentration of the sample. The EMF of the reference electrode is constant. An electronic calculation circuit converts EMF of the sample to the ion concentration of the sample. (Package insert: ISE reagent. Roche Diagnostics; V14.0, 02/2018)

Bicarbonate:

This is a photometric rate reaction. Bicarbonate (HCO3[-]) reacts with phosphoenolpyruvate in the presence of phosphoenolpyruvate carboxylase to produce oxaloacetate and phosphate. The oxaloacetate produced is coupled with reduced nicotinamide adenine dinucleotide (NADH) in the presence of malate dehydrogenase to produce malate and NAD(+). The consumption of NADH causes a decrease in absorbance and is monitored in the ultraviolet range of 320 to 400 nm. The rate of change is directly proportional to the concentration of bicarbonate.(Package insert: Bicarbonate reagent. Roche Diagnostics; 04/2019)

Anion Gap:

This is a calculated result. The following equation is used to calculate the anion gap (A gap): A gap = Na - (Cl + HCO3[-])



Electrolyte Panel, Serum

PDF Report

No

Day(s) Performed

Monday through Sunday

Report Available

Same day/1 to 2 days

Specimen Retention Time

1 week

Performing Laboratory Location

Rochester

Fees & Codes

Fees

- Authorized users can sign in to <u>Test Prices</u> for detailed fee information.
- Clients without access to Test Prices can contact <u>Customer Service</u> 24 hours a day, seven days a week.
- Prospective clients should contact their account representative. For assistance, contact <u>Customer Service</u>.

Test Classification

This test has been cleared, approved, or is exempt by the US Food and Drug Administration and is used per manufacturer's instructions. Performance characteristics were verified by Mayo Clinic in a manner consistent with CLIA requirements.

CPT Code Information

80051-Electrolyte Panel (if all 4 are performed) 82435-Chloride (if all 4 are not performed) 84295- Sodium (if all 4 are not performed) 84132-Potassium (if all 4 are not performed) 82374-Bicarbonate (if all 4 are not performed)

LOINC[®] Information

Test ID	Test Order Name	Order LOINC [®] Value
ELPSR	Electrolyte Panel, S	24326-1
Result ID	Test Result Name	Result LOINC [®] Value
AGAP	Anion Gap	33037-3
CL	Chloride, S	2075-0
HCO3	Bicarbonate, S	1963-8
NAS	Sodium, S	2951-2



Electrolyte Panel, Serum

KS Potassium, S	2823-3
-----------------	--------